



ELSEVIER

European Journal of Cardio-thoracic Surgery xxx (2009) xxx–xxx

 EUROPEAN JOURNAL OF
 CARDIO-THORACIC
 SURGERY

www.elsevier.com/locate/ejcts

Rotation thromboelastometry (ROTEM[®]) stability and reproducibility over time[☆]

Oliver Michel Theusinger^{a,1,*}, Johannes Nürnberg^{a,1}, Lars M. Asmis^b,
 Burkhardt Seifert^c, Donat Rudolph Spahn^a

^a Institute of Anesthesiology, University Hospital Zurich, CH - 8091 Zurich, Switzerland

^b Clinic of Hematology, University Hospital Zurich, Zurich, Switzerland

^c Biostatistics Unit, Institute of Social and Preventive Medicine, University of Zurich, Zurich, Switzerland

Received 22 April 2009; received in revised form 21 July 2009; accepted 29 July 2009

Abstract

Background: Thromboelastometry is a whole blood assay performed to evaluate the viscoelastic properties during blood clot formation and lysis. Rotation thromboelastography (ROTEM[®], Pentapharm GmbH, Munich, Germany) has overcome some of the limitations of classic thromboelastography. So far, no clinical validation on reproducibility (inter- and intra-assay variability) and sample stability over time has been published. **Methods:** To evaluate the pre-analytic aspects, sample stability over time was assessed in 48 patients in eight age groups. Citrated blood was stored at room temperature. Tests measured every 30 min from T 0 min up to T 120 min on two ROTEM[®] devices were INTEM (ellagic acid activated intrinsic pathway), EXTEM (tissue factor-triggered extrinsic pathway) and FIBTEM (with platelet inhibitor (cytochalasin D) evaluating the contribution of fibrinogen to clot formation). Precision by intra- and inter-assay variability was evaluated at two points of time in 10 volunteers. Finally, reference intervals and effect of age and sex were evaluated. **Results:** Blood was stable over 120 min and no significant differences in ROTEM[®] results were found. Maximum clot firmness measurements had a coefficient of variation of <3% for EXTEM, <5% for INTEM and <6% for FIBTEM. For clot formation time, the coefficient of variation was <4% for EXTEM and <3% for INTEM. Coefficient of variation for angle alpha was <3% for EXTEM and <6% for INTEM. The coefficient of variation for clotting time was <15% for both EXTEM and INTEM. Small but significant differences between ROTEM[®] devices were found for maximum clot firmness in FIBTEM and INTEM as well as clot formation time and alpha angle in INTEM. **Conclusions:** ROTEM[®] yields stable results over 120 min with a minimal variability on the same ROTEM[®] device. However, small but significant differences between ROTEM[®] devices were observed. Analysis should be performed on the same ROTEM[®] device if small differences are of importance for treatment.

© 2009 European Association for Cardio-Thoracic Surgery. Published by Elsevier B.V. All rights reserved.

Keywords: Thromboelastometry; Blood coagulation; Haemostasis; Thromboelastography

1. Introduction

Rotation thromboelastometry (ROTEM[®]) (ROTEM[®] delta, Pentapharm GmbH, Munich, Germany), a methodology based on thromboelastography originally described by Hartert more than 50 years ago [1], is frequently used today to rapidly assess the visco-elastic properties of the developing clot in cardiac and transplant surgery as well as following trauma [2–5]. ROTEM[®] documents the interaction of platelets with the coagulation factors from initial platelet–fibrin interaction, through platelet aggregation, clot strengthening and fibrin cross-linking to eventual clot lysis. Within 30 min, a ROTEM[®]

tracing provides information on clotting factor activity, platelet function and any clinically significant fibrinolysis [6,7].

The goal of this study was to assess stability and reproducibility over time of ROTEM[®]. The lack of such a study was already mentioned by Dunning et al. [8] in 2008 and by Samama and Ozier [9] in 2003. Possible changes due to age and sex were also of interest. In thromboelastography (TEG[®] Haemoscope Corporation, Skokie, IL, USA), stability of results over time is only achieved following a 30-min waiting time after blood draw and before analysis [10]. Immediate sample analysis, however, is desirable in the setting of an acute bleeding event. We thus assessed stability over time for ROTEM[®].

2. Material and methods

This clinical trial was performed after obtaining authorisation by the local ethics committee (Kantonale

[☆] *Financial support:* The work carried out in this article was supported by departmental funds; material was provided by Axon Lab AG, 5405 Baden-Dättwil, Switzerland.

* Corresponding author. Tel.: +41 44 255 26 95; fax: +41 44 255 44 09.

E-mail address: oliver.theusinger@usz.ch (O.M. Theusinger).

¹ Both authors contributed equally to this study.

Ethikkommission, Kanton Zurich, Switzerland, Study number StV 27-2007).

Sample size ($n = 48$) was chosen based on statistical considerations and on reported general practice. Horn and Pesce showed that a minimum of 39 patients is needed for establishing a 95% reference interval by power analysis [11]. In addition, Friedberg et al. reported in 2007 [12], that 50% of laboratories enrol 21–50 subjects when establishing reference intervals. As a possible age or gender effect was of interest, three male and three female patients in each of the following eight age categories were included: below 20 years; 20–30 years; 30–40 years; 40–50 years; 50–60 years; 60–70 years; 70–80 years; and above 80 years.

Inclusion criteria were scheduled for a non-emergent operation and signed written informed consent. Exclusion criteria were known malignancy or immunosuppression, known coagulation disorders, anticoagulation in any form, current treatment with heparin (other than routine pre-operative thromboembolic prophylaxis with 3000 IU low-molecular-weight heparin administered subcutaneously the evening prior to the operation), use of acetyl salicylic acid within the past 5 days, use of non-steroidal anti-inflammatory agents within the past 24 h, known renal diseases or plasma concentration of creatinine more than 120 mM and liver diseases or increased plasma concentration of aspartate aminotransferase ($>50 \text{ U l}^{-1}$) or alanine aminotransferase ($>50 \text{ U l}^{-1}$) as well as patients not capable of understanding the German language.

The study's main objectives were to validate ROTEM[®] by investigating (1) pre-analytic aspects (sample stability), (2) reproducibility and precision of ROTEM[®] (intra- and inter-assay variability), and (3) reference intervals and effect of age and sex.

Several predefined tests were assessed: INTEM (ellagic acid activated intrinsic pathway), EXTEM (tissue factor triggered extrinsic pathway) and FIBTEM (with platelet inhibitor (cytochalasin D) evaluating the contribution of fibrinogen to clot formation). These three assays are performed in citrated samples and represent counterparts of routine tests of plasmatic coagulation: fibrinogen, prothrombin time and activated partial thromboplastin time. The two main differences between thromboelastometric tests and plasmatic coagulation tests are (i) the former is performed on whole blood, the latter on plasma, and (ii) the former measures processes involving thrombin generation, clot formation and clot lysis, the latter processes leading up and until the initial generation of thrombin. By using inhibitors of platelet function, the developers of the assay have derived a functional test of fibrinogen, called FIBTEM. This can be correlated with functional fibrinogen tests in plasma. By comparing EXTEM, the thromboelastometric test that uses tissue factor (comparable to the prothrombin time test) as an activator, with FIBTEM, one can deduce relevant information regarding platelet function and factor XIII function, as both of these contribute to EXTEM measurement.

In EXTEM, the extrinsic pathway is activated by thromboplastin from rabbit brain to assess clot formation and fibrinolysis. In INTEM, the intrinsic pathway is activated by a contact activator to assess the clot formation and fibrin polymerisation. In FIBTEM, the extrinsic pathway is activated

by tissue factor in presence of a platelet inhibitor to assess the functional fibrinogen level.


Ten volunteers were sufficient to investigate precision and reproducibility according to power analysis with 95% confidence interval (CI). To investigate precision and reproducibility of ROTEM[®] analysis, for inter-assay (on one single ROTEM[®] device), reproducibility testing was performed in the volunteers, after having obtained written informed consent, withdrawing three tubes (Vacutainer Brand, Belliver Industrial Estate, Plymouth, UK, 4.5 ml, 9 NC 0.129 M, a total of 18 ml per volunteer) of citrated blood, the first tube (Vacutainer Brand, Belliver Industrial Estate, Plymouth, UK, 4.5 ml, 9 NC 0.129 M) drawn was discarded to exclude coagulation activation due to vein puncture and blood withdrawal. Inter-device variability was tested using the same samples on a second ROTEM[®] device. Intra-assay variability was tested by performing duplicate measurements at the time of the first blood withdrawal on the same device using different channels and calculating the coefficient of variation. At a second point of time, that is, 1 week later, another three tubes (Vacutainer Brand, Belliver Industrial Estate, Plymouth, UK, 4.5 ml, 9 NC 0.129 M, a total of 18 ml per volunteer) of citrated blood were again withdrawn from the same 10 volunteers to assess the reproducibility of the two ROTEM[®] devices at the second point of time (week 2) as compared with the first week. The 10 volunteers were nurses or physicians from the Institute of Anaesthesiology, of the University Hospital of Zurich, and the tests ran for 30 min.

After obtaining written informed consent from the patients, eight tubes (Vacutainer Brand, Belliver Industrial Estate, Plymouth, UK, 4.5 ml, 9 NC 0.129 M) of citrated blood were withdrawn, a total of 36 ml of blood per patient. The first tube (Vacutainer Brand, Belliver Industrial Estate, Plymouth, UK, 4.5 ml, 9 NC 0.129 M) drawn was discarded to exclude coagulation activation due to vein puncture and blood withdrawal. In the 48 patients, INTEM, EXTEM and FIBTEM were performed at $T_1 = 0 \text{ min}$, $T_2 = 30 \text{ min}$, $T_3 = 60 \text{ min}$, $T_4 = 90 \text{ min}$ and $T_5 = 120 \text{ min}$. The blood was stored at room temperature. Tests were performed at 37°C . Measurements for this test series ran for 60 min.

2.1. Parameters of rotation thromboelastometry

Q1

ROTEM[®] defines various parameters to describe the dynamics, the size and the firmness of clot during clot formation and lysis (Fig. 1). The clotting time (CT) is the period from the start of the analysis until the start of clot formation, normally until the 2-mm amplitude is reached. The clot formation time (CFT) is defined as the period until an amplitude of 20 mm is reached. The angle alpha is given by the angle between the centre line and a tangent to the curve through the 2-mm amplitude point. The maximum amplitude of the curve is defined as the maximum clot firmness (MCF). The amplitude at different points of time is described by A5 till A30, whereby the number refers to the time since the start of the test. The clot lysis index (CLI) at 30 and 60 min (CLI30, CLI60) describes the ratio between the maximum clot firmness and the amplitude 30 and 60 min after clotting time, and gives information about the fibrinolysis. The maximum lysis (ML) represents the maximum fibrinolysis detected during the measurement.



Q2 Fig. 1. Modified, with authorisation of Pentapharm GmbH, Munich, Germany. CT (clotting time): time from the start of measurement until initiation of clotting – initiation of clotting, thrombin formation, start of clot polymerisation. CFT (clot formation time): time from initiation of clotting until a clot firmness of 20 mm is detected – fibrinpolymerisation, stabilisation of the clot with thrombocytes and F XIII. MCF (maximum clot firmness): firmness of the clot – increased stability of the clot by the polymerised fibrin, thrombocytes as well as F XIII. Lysis: reduction of the clot firmness after MCF in relation to MCF – stability of the clot. A maximum lysis (ML) <15% is considered normal, a ML >15% within 1 h is indicative of exaggerated fibrinolysis (hyperfibrinolysis).

summarised as mean \pm SD and median with CI where appropriate. Analysis of variance (ANOVA) for repeated measures was used to analyse all parameters with *post hoc* comparison and Bonferroni correction. For age and gender, regression analyses were used. *p* Values of 0.05 or less are considered significant.

3. Results

3.1. Baseline information

The 48 patients had a mean age of 50 ± 22 years, ranging from 17 to 87 years. For the 24 women, the mean age was 50 ± 22 years, ranging from 17 to 85 years. For the 24 men, the mean age was 50 ± 22 years, ranging from 17 to 87 years. The 10 volunteers were five men and five women; the overall mean age was 37 ± 14 years, ranging from 20 to 63 years.

3.2. Pre-analytic aspects' stability over time

Over the time of 120 min, there was no significant difference between any parameter in EXTEM, INTEM and FIBTEM, indicating that ROTEM[®] measurements are stable over 2 h at room temperature (Table 1).

3.3. Reproducibility and precision of ROTEM[®] devices (intra- and inter-assay variability)

The reproducibility of results on two ROTEM[®] devices was tested with 10 volunteers. Tests were performed twice with a week's interval in between; and an overall analysis was performed between the two points of time (week 1 vs week 2) and the two ROTEM[®] devices.

The reproducibility of maximum clot firmness in the 10 volunteers at the two points of time in the two ROTEM[®] devices showed no significant difference between the two points of time ($p > 0.200$) and there was no influence by the ROTEM[®] device ($p > 0.200$) in the overall effect for EXTEM (mean 63.0 ± 5.5 mm, reference range 53–72 mm, mean difference -0.1 mm, 95% CI -0.8 to 0.7 mm, maximum difference 1.3%).

For the maximum clot firmness of FIBTEM (mean 11.9 ± 3.7 mm, reference range 9–25 mm, mean difference 0.6 mm, 95% CI 0.2 – 1.0 mm, maximum difference 8.4%) and INTEM (mean 61.6 ± 4.3 mm, reference range 53–72 mm, mean difference 1.7 mm, 95% CI 1.2 – 2.2 mm, maximum difference 3.6%), a significant difference between devices (FIBTEM $p = 0.005$, INTEM $p < 0.001$) but not for the points of time (FIBTEM $p > 0.200$, INTEM $p = 0.187$) was found in the overall analysis (Table 2, Fig. 2).

For EXTEM and FIBTEM, no significant difference between devices or time points was found for clot formation time, clotting time and angle alpha (Table 2). The clot formation time ($p = 0.001$, mean 74.5 ± 20.1 s, reference range 35–110 s, mean difference -7.1 s, 95% CI -10.4 to -3.7 s, maximum difference 14.0%) and angle alpha ($p = 0.013$, mean $75.2 \pm 3.7^\circ$, reference range 70 – 83° , mean difference 1.2° , 95% CI 0.3 – 2.1° , maximum difference 2.8%) were significantly different between devices in INTEM, but no difference for the two points of time was found. For the two

181 The parameters measured in this study were: clotting
182 time, clot formation time, maximum clot firmness, angle
183 alpha, and amplitude after 5, 10, 15, 20, 25 and 30 min.
184

2.2. Test procedure

185 All ROTEM[®] devices used in this study were new and set up
186 by a representative of the local distributor. The tests were
187 performed using the automated pipette programmes accord-
188 ing to the instructions of the manufacturer. For each
189 measurement, a new pin was positioned on the axis of the
190 measurement channel and a new cup was put into the special
191 cup holder of the device. According to the pipetting
192 programme, 20 μ l re-calcification reagent (200 mmol l⁻¹
193 calcium chloride solution) and 20 μ l of the respective
194 activation reagent (FIBTEM: cytochalasin D, EXTEM: throm-
195 boplastin from rabbit brain and INTEM: partial thromboplas-
196 tin phospholipid made of rabbit brain (chloroform extract),
197 ellagic acid) were added into the pre-warmed cup. Then,
198 300 μ l of citrated whole blood was added to the cup and,
199 after a semi-automated mixing step, the cup holder was
200 placed in the measuring position of the ROTEM[®] device. The
201 measurement started automatically when blood was added
202 to the cup and was stopped after 30 or 60 min according to
203 the protocol (EXTEM lot 41194401, INTEM lot 41166301,
204 FIBTEM lot 41147601, star-TEM lot 41166101).
205

2.3. Statistical analyses

206 A trial database within Excel (Microsoft Office 2003,
207 Microsoft Corporation Redmond, WA, USA) was used to store
208 study data (transferred from the ROTEM[®] devices). The
209 statistical analyses were performed using SPSS[®] (version 13,
210 SPSS Inc., Chicago, IL, USA). Continuous variables are
211

Table 1
Q3 Sample stability over the time, mean values and standard deviation.

	CT (s)					CFT (s)				
	T 0	T 30 min	T 60 min	T 90 min	T 120 min	T 0	T 30 min	T 60 min	T 90 min	T 120 min
INTEM										
Mean ± SD	152.3 ± 20.6	153.6 ± 16.9	154.7 ± 20.0	156.5 ± 21.4	156.0 ± 20.1	64.7 ± 15.1	62.9 ± 15.4	64.1 ± 16.1	61.3 ± 15.5	61.6 ± 14.8
EXTEM										
Mean ± SD	55.3 ± 6.4	56.0 ± 6.1	56.9 ± 7.3	56.0 ± 6.3	55.0 ± 6.4	80.3 ± 20.7	79.0 ± 18.6	80.1 ± 21.4	77.3 ± 20.0	77.9 ± 19.2
FIBTEM										
Mean ± SD	53.5 ± 7.7	52.9 ± 5.4	53.0 ± 6.9	52.2 ± 5.6	53.7 ± 11.4	585.6 ± 553.9	696.7 ± 900.2	592.4 ± 847.2	506.4 ± 457.8	817.6 ± 941.7
	Alpha (°)					MCF (mm)				
	T 0	T 30 min	T 60 min	T 90 min	T 120 min	T 0	T 30 min	T 60 min	T 90 min	T 120 min
INTEM										
Mean ± SD	77.1 ± 2.8	77.5 ± 2.9	77.2 ± 2.9	77.7 ± 2.8	77.6 ± 2.9	63.5 ± 5.2	64.5 ± 5.0	63.8 ± 5.2	64.4 ± 5.1	64.0 ± 5.1
EXTEM										
Mean ± SD	73.9 ± 4.1	74.0 ± 3.8	74.0 ± 4.2	74.4 ± 3.9	74.2 ± 3.8	66.7 ± 5.5	66.3 ± 5.0	66.2 ± 5.3	66.3 ± 5.1	66.5 ± 5.3
FIBTEM										
Mean ± SD	69.9 ± 6.9	71.2 ± 5.3	70.8 ± 6.3	70.1 ± 6.1	71.8 ± 5.0	16.1 ± 5.9	16.3 ± 5.8	16.6 ± 6.0	16.7 ± 6.2	16.1 ± 5.6

MCF: maximum clot firmness, CFT; clot formation time, CT; clotting time.

Table 2

MCF: maximum clot firmness, CFT: clot formation time, CT: clotting time and angle alpha changes between two ROTEM® devices and weeks and 2, mean values and standard deviations.

Test	Week	Device	N	MCF (mm) mean ± SD	CFT (s) mean ± SD	CT (s) mean ± SD	Angle alpha mean ± SD
EXTEM	1	1	10	63.3 ± 5.5	104.5 ± 29.1	56.7 ± 13.0	69.4 ± 5.4
EXTEM	1	2	10	63.1 ± 5.8	98.3 ± 31.1	61.0 ± 11.5	70.5 ± 5.8
EXTEM	2	1	10	62.7 ± 5.5	99.3 ± 24.6	59.0 ± 6.5	70.3 ± 4.7
EXTEM	2	2	10	62.8 ± 5.3	96.8 ± 25.1	58.9 ± 4.5	70.6 ± 4.9
FIBTEM	1	1	10	11.4 ± 4.0*	n.a.	57.9 ± 10.4	n.a.
FIBTEM	1	2	10	12.2 ± 4.2*	n.a.	57.3 ± 7.3	n.a.
FIBTEM	2	1	10	11.7 ± 3.2*	n.a.	54.6 ± 7.2	n.a.
FIBTEM	2	2	10	12.1 ± 3.2*	n.a.	54.1 ± 5.2	n.a.
INTEM	1	1	10	60.9 ± 4.8*	77.2 ± 22.5*	155.7 ± 17.7**	74.8 ± 4.0*
INTEM	1	2	10	62.9 ± 4.3*	68.8 ± 21.1*	157.0 ± 12.0**	76.1 ± 4.0*
INTEM	2	1	10	60.5 ± 4.0*	78.8 ± 18.6*	164.1 ± 12.6**	74.4 ± 3.4*
INTEM	2	2	10	61.9 ± 4.0*	73.1 ± 18.0*	160.3 ± 18.7**	75.5 ± 3.4*

* $p < 0.050$ for ROTEM® devices.

** $p < 0.050$ for the weeks.

points of time, a significant difference was only found for the clotting time (mean week 1 156.4 ± 14.9 s, mean week 2 162.2 ± 15.7 s, reference range 100–240 s) in INTEM ($p = 0.046$). For FIBTEM, the only analysis that is of importance is maximum clot firmness; all other values are not to be used in the interpretation of FIBTEM.

Reproducibility of results for maximum clot firmness was 97% in EXTEM, 95% in INTEM and 94% in FIBTEM. The coefficient of variation was calculated in the first point of time (week 1) on the same device. Coefficient of variation for maximum clot firmness in EXTEM was <3%, INTEM <5% and FIBTEM <6%. For clot formation time, the coefficient of variation was <4% for EXTEM and <3% INTEM. Coefficient of variation for the angle alpha was <3% for EXTEM and <6% for INTEM. Clotting time had a coefficient of variation of <15% for EXTEM and INTEM.

3.4. Early prediction of maximum clot firmness

As tests for the 48 patients exceeded 60 min, we calculated for each point of time the percentage of the

final maximum clot firmness, showing that in an overall analysis, after 10 min of running time, maximum clot firmness in EXTEM, INTEM and FIBTEM reached at least 98% of the final maximum clot firmness value.

3.5. Reference intervals

Reference intervals were calculated by means, according to international guidelines, and adding one standard deviation on either side. We also calculated reference intervals by gender (Table 3). We were able to show that reference intervals calculated in this study were closer together than those set by the manufacturer. A gender-specific influence is also to be seen.

3.6. Effect of age and gender

With advancing age, maximum clot firmness increased significantly in all tests ($p < 0.001$) by a mean of 0.1 mm per year of age over 20 years. Angle alpha also increased significantly for all tests with a mean of 0.07° per year of age




Fig. 2. Boxplot of MCF: maximum clot firmness for all tests, representing median by device 1 and 2 and point of time 1 and 2, * $p < 0.050$ for differences between ROTEM[®] devices.

Table 3

Reference values by the manufacture, reference values calculated in this study and calculated by gender in this study.

	Manufacturer	Overall	Male	Female
EXTEM MCF (mm)	50–70	61–72	59–72	64–72
INTEM MCF (mm)	50–72	59–68	56–68	61–69
FIBTEM MCF (mm)	9–25	10–22	8–22	13–21
EXTEM CFT (s)	34–159	61–97	62–108	59–86
INTEM CFT (s)	30–110	49–77	51–86	47–68
FIBTEM CFT (s)	n.a.	n.a.	n.a.	n.a.
EXTEM CT (s)	38–79	48–63	50–65	46–62
INTEM CT (s)	100–240	133–177	131–182	134–171
FIBTEM CT (s)	43–75	48–58	51–59	46–56
EXTEM angle alpha (°)	63–83	70–78	68–78	73–78
INTEM angle alpha (°)	70–83	75–80	73–80	76–81
FIBTEM angle alpha (°)	n.a.	n.a.	n.a.	n.a.

MCF: maximum clot firmness, CFT: clot formation time, CT: clotting time.

over 20 years in EXTEM ($p = 0.009$), 0.14° in FIBTEM ($p = 0.007$) and 0.04° in INTEM ($p = 0.019$). Clot formation time decreased significantly in EXTEM ($p = 0.008$) and INTEM ($p = 0.014$), with a mean of -0.4 s per year of age over 20

Table 4

Sex related differences in the main parameters of ROTEM[®], mean values and standard deviations.

	Age (years)	CT (s)	CFT (s)	Alpha angle (°)	MCF (mm)
INTEM					
Females ($n = 24$)	50.8 ± 21.9	152.5 ± 18.1	$57.7 \pm 10.5^*$	$78.4 \pm 2.1^*$	$64.8 \pm 3.8^*$
Males ($n = 24$)	52.0 ± 22.1	156.8 ± 25.5	$68.1 \pm 17.5^*$	$76.5 \pm 3.2^*$	$62.2 \pm 6.1^*$
EXTEM					
Females ($n = 24$)	50.8 ± 21.9	53.9 ± 7.8	$72.6 \pm 13.3^*$	$75.3 \pm 2.7^*$	$68.1 \pm 4.2^*$
Males ($n = 24$)	52.0 ± 22.1	57.7 ± 7.3	$85.2 \pm 23.2^*$	$72.9 \pm 4.6^*$	$65.3 \pm 6.4^*$
FIBTEM					
Females ($n = 24$)	50.8 ± 21.9	51.1 ± 5.0	n.d.	71.0 ± 5.6	17.3 ± 4.1
Males ($n = 24$)	52.0 ± 22.1	55.0 ± 4.2	n.d.	70.4 ± 6.5	15.0 ± 7.2

MCF: maximum clot firmness, CFT: clot formation time, CT: clotting time.

* $p < 0.050$ comparison male versus female.

years in EXTEM and -0.2 s per year of age over 20 years in INTEM. The age had no influence on clotting time in all tests.

Maximum clot firmness ($p < 0.001$ for EXTEM and INTEM) and angle alpha ($p = 0.009$ in EXTEM, $p = 0.019$ in INTEM) were significantly higher and clot formation time lower ($p = 0.008$ in EXTEM, $p = 0.014$ in INTEM) in EXTEM and INTEM in women than in men, indicating a somewhat greater coagulability in women. Similar changes in FIBTEM (towards hypercoagulability) did not reach statistical significance. Clotting time was similar in all age groups and in both sexes (Table 4).

4. Discussion

Our analysis of pre-analytic aspects showed that ROTEM[®] tests yield stable results from citrated blood that was recalcified during the first 120 min. This is of high practical relevance since the exact delay from blood drawing to testing would be difficult to standardise, particularly in major trauma. This is in contrast to the thromboelastography, where a previous report showed that parameters are unstable in the first 30 min after blood withdrawal [10].

Evaluation of the second aim (reproducibility: inter-assay, same device) showed a good reproducibility of results on the same ROTEM[®] device as well as over a time period of 120 min. Statistically significant differences on different ROTEM[®] devices occurred for the maximum clot firmness of FIBTEM and INTEM as well as the clot formation time and alpha angle of INTEM as mentioned above. For the interpretation of INTEM, these results are statistically significant but not of any clinical relevance because the values are largely within the reference values and would not be implemented in a treatment. Reference intervals calculated for our study and for gender were found to be within a narrower interval than those provided by the manufacturer. According to the manufacturer, the company's reference values are for orientation purposes and should be validated individually, as they may vary from lab to lab, depending on blood sampling technology and other pre-analytical factors.

Evaluation of the third aim (precision: inter-device) demonstrated reproducibility of ROTEM[®] measurements over time and when the blood sample is analysed on the same ROTEM[®] device sequentially. In addition, citrated blood samples analysed after 0–120 min yield similar results in all

ROTEM[®] tests assessed. However, when a blood sample is analysed on different ROTEM[®] devices simultaneously, statistically significant differences ($p < 0.050$) were detected in the overall comparison for the maximum clot firmness in FIBTEM (maximum difference 8.4%) and in INTEM (maximum difference 3.6%) as well as for clot formation time (maximum difference 14%) and angle alpha (maximum difference 2.8%) in INTEM. In clinical use, when using algorithms, a difference in maximum clot firmness of 1–2 mm in FIBTEM may result in a premature or delayed treatment of the patient. Therefore, when working with exact algorithms, the blood of one individual patient should be measured on one single ROTEM[®] device.

Compared with the results of Lang et al. [13], we have found a somewhat lower variability for the maximum clot firmness in the FIBTEM test (coefficient of variability for maximum clot firmness <6%). This may be due to the fact that, in our study, only three persons performed all ROTEM[®] tests, which is impossible when pooling data from different centres. Performing ROTEM[®] tests by persons familiar with the ROTEM[®] technology thus appear to reduce the variability.

A progressive change of the ROTEM[®] parameters towards hypercoagulability with advancing age as observed in this study has also been described by Ng [14] for thromboelastography (TEG[®] Haemoscope Corporation, Skokie, IL, USA), and is in keeping with previous reports describing hypercoagulability in elderly people [14,15]. Since the ROTEM[®] parameters measured in the current study were still within the reference ranges provided by the manufacturer, an age-specific adaptation of the reference ranges may not be required for ROTEM[®]. Nevertheless, the trend of increased clot firmness at advanced age should be kept in mind in the clinical interpretation of ROTEM[®] results.

We found differences between men and women in ROTEM[®] parameters, which indicate a faster development of the clot and greater clot strength in women as compared with men (Table 4). Unfortunately, no measurements of fibrinogen according to Clauss were performed in this study. Nevertheless, this may represent a factor predisposing women to thrombotic complications [16].

A limitation of this study is that we did not take into account the effective temperature in patients since it is known to impair results and interpretation of ROTEM[®] [17]. We think that this point is of limited importance since the ROTEM[®] devices work at a temperature of 37 °C and the 300 µl of blood drawn are at this same temperature and the patients were not hypothermic at the moment the blood was drawn.

ROTEM[®] results cannot give exact recommendations on the amount of blood products or coagulation factors to be administered. However, ROTEM[®] can guide the clinician as to which type of treatment may be most helpful to treat coagulopathy during surgery or in trauma. In 2007, Rugeri et al. demonstrated in trauma patients that ROTEM[®] can rapidly detect trauma-related coagulopathy and might be helpful in guiding treatment [18]. Theusinger et al. [19] published a ROTEM[®]-based transfusion algorithm to guide individual goal-directed transfusions. A retrospective analysis by Anderson et al. [20] in 990 patients showed that the use of ROTEM[®] significantly decreases the use of red blood cells

and other blood products after cardiac surgery. This is particularly important given the fact that red blood cell transfusion in cardiac surgery is consistently associated with significant morbidity and increased mortality [21–24]. Last but not least, Spalding et al. have demonstrated that a ROTEM[®]-based coagulation algorithm decreased total transfusion costs in cardiac surgery [25].

5. Conclusion

ROTEM[®] measurements of EXTEM, INTEM and FIBTEM are reproducible and stable over time, regardless of delay from blood withdrawal to analysis (range 0–120 min after blood withdrawal). There is a high reproducibility with the coefficient of variation <6% in all assays. Coefficient of variation for maximum clot firmness in EXTEM was <3%, INTEM <5% and FIBTEM <6%; they are small but statistically significant. With age, there is a tendency of hypercoagulability and also with sex: women seem to coagulate better than men but within range. To avoid problems in treating coagulopathies when working with an algorithm or to interpret the evolution of ROTEM[®] parameters over time, ROTEM[®] tests should preferably be performed on the same ROTEM[®] device, because in our study one ROTEM[®] device yielded slightly but statistically significantly different results for maximum clot firmness and clot formation time.

References

- [1] Hartert H. Blutgerinnungsstudien mit der Thrombelastographie, ein neues Untersuchungsverfahren. *Klin Wochenschr* 1948;577–83.
- [2] Samama CM. Thromboelastography: the next step. *Anesth Analg* 2001;92:563–4.
- [3] Vig S, Chitolie A, Bevan DH, Halliday A, Dormandy J. Thromboelastography: a reliable test? *Blood Coagul Fibrinolysis* 2001;12:555–61.
- [4] Mallett SV, Cox DJ. Thrombelastography. *Br J Anaesth* 1992;69:307–13.
- [5] Di Benedetto P, Baciarello M, Cabetti L, Martucci M, Chiaschi A, Bertini L. Thrombelastography. Present and future perspectives in clinical practice. *Minerva Anestesiol* 2003;69:501–9 [509–15].
- [6] Artang R, Jensen E, Pedersen F, Frandsen NJ. Thrombelastography in healthy volunteers, patients with stable angina and acute chest pain. *Thromb Res* 2000;97:499–503.
- [7] Ganter MT, Hofer CK. Coagulation monitoring: current techniques and clinical use of viscoelastic point-of-care coagulation devices. *Anesth Analg* 2008;106:1366–75.
- [8] Dunning J, Versteegh M, Fabbri A, Pavie A, Kolh P, Lockowandt U, Nashef SA. Guideline on antiplatelet and anticoagulation management in cardiac surgery. *Eur J Cardiothorac Surg* 2008;34:73–92.
- [9] Samama CM, Ozier Y. Near-patient testing of haemostasis in the operating theatre: an approach to appropriate use of blood in surgery. *Vox Sang* 2003;84:251–5.
- [10] Camenzind V, Bombeli T, Seifert B, Jamnicki M, Popovic D, Pasch T, Spahn DR. Citrate storage affects thrombelastograph analysis. *Anesthesiology* 2000;92:1242–9.
- [11] Horn PS, Pesce AJ. Reference intervals: an update. *Clin Chim Acta* 2003;334:5–23.
- [12] Friedberg RC, Souers R, Wagar EA, Stankovic AK, Valenstein PN. The origin of reference intervals. *Arch Pathol Lab Med* 2007;131:348–57.
- [13] Lang T, Bauters A, Braun SL, Potzsch B, von Pape KW, Kolde HJ, Lakner M. Multi-centre investigation on reference ranges for ROTEM thromboelastometry. *Blood Coagul Fibrinolysis* 2005;16:301–10.
- [14] Ng KF. Changes in thrombelastograph variables associated with aging. *Anesth Analg* 2004;99:449–54 [Table of contents].
- [15] Javorschi S, Richard-Harston S, Labrousche S, Manciet G, Freyburger G. Relative influence of age and thrombotic history on hemostatic parameters. *Thromb Res* 1998;91:241–8.

- 469
471
472
473
474
475
476
477
478
479
480
481
482
483
484
485
486
487
488
- [16] Grassi M, Kisiailiou A, Assanelli D, Mozzini C, Archetti S, Pezzini A. Evidence of major genes effects on serum homocysteine and fibrinogen levels, and premature ischemic heart disease in Italian extended families. *Hum Hered* 2008;66:50–60.
- [17] Dirkmann D, Hanke AA, Gorlinger K, Peters J. Hypothermia and acidosis synergistically impair coagulation in human whole blood. *Anesth Analg* 2008;106:1627–32.
- [18] Rugeri L, Levrat A, David JS, Delecroix E, Floccard B, Gros A, Allaouchiche B, Negrier C. Diagnosis of early coagulation abnormalities in trauma patients by rotation thrombelastography. *J Thromb Haemost* 2007;5: 289–95.
- [19] Theusinger OM, Spahn DR, Ganter MT. Transfusion in trauma: why and how should we change our current practice? *Curr Opin Anaesth* 2009;22:305–12.
- [20] Anderson L, Quasim I, Soutar R, Steven M, Macfie A, Korte W. An audit of red cell and blood product use after the institution of thromboelastometry in a cardiac intensive care unit. *Transfus Med* 2006;16:31–9.
- [21] Karkouti K, Wijeyesundera DN, Yau TM, Callum JL, Cheng DC, Crowther M, Dupuis JY, Fremes SE, Kent B, Laflamme C, Lamy A, Legare JF, Mazer CD, McCluskey SA, Rubens FD, Sawchuk C, Beattie WS. Acute kidney injury after cardiac surgery: focus on modifiable risk factors. *Circulation* 2009;119:495–502.
- [22] Kulier A, Levin J, Moser R, Rumpold-Seitlinger G, Tudor IC, Snyder-Ramos SA, Moehnle P, Mangano DT. Impact of preoperative anemia on outcome in patients undergoing coronary artery bypass graft surgery. *Circulation* 2007;116:471–9.
- [23] Murphy GJ, Reeves BC, Rogers CA, Rizvi SI, Culliford L, Angelini GD. Increased mortality, postoperative morbidity, and cost after red blood cell transfusion in patients having cardiac surgery. *Circulation* 2007;116: 2544–52.
- [24] Surgenor SD, Kramer RS, Olmstead EM, Ross CS, Sellke FW, Likosky DS, Marrin CA, Helm Jr RE, Leavitt BJ, Morton JR, Charlesworth DC, Clough RA, Hernandez F, Frumiento C, Benak A, DioData C, O'Connor GT. The association of perioperative red blood cell transfusions and decreased long-term survival after cardiac surgery. *Anesth Analg* 2009;108:1741–6.
- [25] Spalding GJ, Hartrumpf M, Sierig T, Oesberg N, Kirschke CG, Albes JM. Cost reduction of perioperative coagulation management in cardiac surgery: value of "bedside" thrombelastography (ROTEM). *Eur J Cardiothorac Surg* 2007;31:1052–7.

488
489
490
491
492
493
494
495
496
497
498
499
500
501
502
503
504
505
506
507